GEFRAN

TPSADA

DIGITAL AUTOZERO & SPAN OPTION



Main Features

- · Coarse Magnetic Autozero
- · Fine Magnetic Autozero
- Calibration Signal 80% FS
- Magnetic Autospan
- Partial reset function
- Total reset function

Digital "Autozero & Span" option makes the pressure transmitter TPSA Series a product extremely innovative and versatile.

Designed thanks to the advanced electronics used by TPSA, this function allows zero and span adjustment in a simple, fast and reliable way, avoiding to use analogue trimmer, potential weak point and source of possible drifts or failures.


The starting is user friendly, simply touching the specific target marked on the sensor body with a pen with magnetic head (included as

Within few seconds the operation is concluded, without the need to dismantle or to open the instrument.

TECHNICAL DATA

Autozero Setting Time Autozero Setting Time Fine Autozero Adjustment Fine Autozero Adjustment Amplitude Calibration Function Calibration Function Setting Time Autospan Setting Time Autospan Setting Time Partial Reset Partial Reset Setting Time 110 seconds Resolution 6 mV (voltage output); 12 μA (current output) ±100 mV (voltage output), ±0.16 mA (current output) by successive steps with maximum setting time 5 sec. per step Signal output generation of 80% FS @ °T Amb. >1 sec. (by contacts closed in CAL position) ±5%FS max with span setting within the sensor accuracy class, @ °T Amb. 110 sec. (by contacts closed in CAL position) Restore of zero factory setting Partial Reset Setting Time 3060 sec		
Fine Autozero Adjustment Fine Autozero Adjustment Amplitude ±100 mV (voltage output), ±0.16 mA (current output) by successive steps with maximum setting time 5 sec. per step Calibration Function Signal output generation of 80% FS @ °T Amb. Calibration Function Setting Time >1 sec. (by contacts closed in CAL position) Autospan 45% FS max with span setting within the sensor accuracy class, @ °T Amb. Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Autozero	±10%FS max with zero setting within the sensor accuracy class, @ °T Amb.
Fine Autozero Adjustment Amplitude ±100 mV (voltage output), ±0.16 mA (current output) by successive steps with maximum setting time 5 sec. per step Calibration Function Signal output generation of 80% FS @ °T Amb. Calibration Function Setting Time >1 sec. (by contacts closed in CAL position) Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Autozero Setting Time	110 seconds
setting time 5 sec. per step Calibration Function Signal output generation of 80% FS @ °T Amb. Calibration Function Setting Time >1 sec. (by contacts closed in CAL position) 4utospan 45% FS max with span setting within the sensor accuracy class, @ °T Amb. Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Fine Autozero Adjustment	Resolution 6 mV (voltage output); 12 μ A (current output)
Calibration Function Signal output generation of 80% FS @ °T Amb. Calibration Function Setting Time >1 sec. (by contacts closed in CAL position) 45%FS max with span setting within the sensor accuracy class, @ °T Amb. Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Fine Autozero Adjustment Amplitude	±100 mV (voltage output), ±0.16 mA (current output) by successive steps with maximum
Calibration Function Setting Time >1 sec. (by contacts closed in CAL position) Autospan ±5%FS max with span setting within the sensor accuracy class, @ °T Amb. Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting		setting time 5 sec. per step
Autospan ±5%FS max with span setting within the sensor accuracy class, @ °T Amb. Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Calibration Function	Signal output generation of 80% FS @ °T Amb.
Autospan Setting Time 110 sec. (by contacts closed in CAL position) Partial Reset Restore of zero factory setting	Calibration Function Setting Time	>1 sec. (by contacts closed in CAL position)
Partial Reset Restore of zero factory setting	Autospan	±5%FS max with span setting within the sensor accuracy class, @ °T Amb.
The state of the s	Autospan Setting Time	110 sec. (by contacts closed in CAL position)
Partial Reset Setting Time 3060 sec	Partial Reset	Restore of zero factory setting
	Partial Reset Setting Time	3060 sec
Total Reset Restore of complete factory setting	Total Reset	Restore of complete factory setting
Total Reset Setting Time >60 sec	Total Reset Setting Time	>60 sec
Function Activation By pen with magnetic head (PKIT 312)	Function Activation	By pen with magnetic head (PKIT 312)

INSTALLATION DRAWINGS



TPSADA, due to the autozero & span board, is 5 mm (0.197 in) longer, so dimension "L" on TK data sheet **Dimensions:**

becames 75 mm (2.95 in).

CAL function pin: The activation of CAL function (Calibration and Autospan) is made by closing the contact between the

pin E-F - 6 pole connector version (V) following pin:

- 7 pole connector version (P) pin 5-6

Please refer to the standard data sheet for all the remaining mechanical and electrical specifications.

ORDERING INFORMATION Pressure transmitter TPSADA **OUTPUT SIGNAL** Mechanical and/or electrical cha-Standard racteristics differing from standard 0,1 ... 10,1 Vdc may be arranged on request. С 4...20 mA Ε 0...10 Vdc Ν **RESPONSE TIME** On request Fast (< 1 msec) 0,1 ... 5,1 Vdc 0 ... 5 Vdc M **ACCURACY** Р 1 ... 5 Vdc ± 0,1% FSO typical 1 ... 10 Vdc Q ≥100bar/1500psi 1 ... 6 Vdc R Т ±0,15% FSO typical PRESSURE CONNECTION <100bar/1500psi Standard G 1/4 gas male (DIN 3852-A) **MEASUREMENT RANGE** On request bar psi 7/16-20 UNF-2A male 2 B05D 0..50 P75D 0..750 (SAE 4 per AS4395-E) B01C 0..100 P₁₅C 0..1500 G 1/2A (DIN 16288) 3 P₂₀C B16D 0..160 0..2000 G 1/4 gas female 4 B₀2C 0..200 P25C 0..2500 1/8-27 NPT female 5 B25D 0..250 **P03M** 0..3000 1/4 - 18 NPT female 6 1/4 - 18 NPT male B35D 0..350 **P05M** 0..5000 7 B04C 0..400 P75C 0..7500 M14 x 1,5 male 8 1/8 - 27 NPT male 9 B05C 0..500 P10M 0..10000 P15M 0..600 G 1/4 gas male (DIN 3852-E) Ε B06C 0..15000 B07C 0..700 M12 x 1,5 male R B01M 0..1000 7/16-20 UNF-2A male Κ (SAE 4for J1926-2) (*) 7/16-20 UNF-2A female CALIBRATION STANDARDS F (SAE 4) Instruments manufactured by Gefran (*) Max. working pressure: 630 bar are calibrated against precision pres-(9137 psi) sure calibration equipment wich is tra-**ELECTRICAL CONNECTION** ceable to International Standards 4-pole connector solenoid | E (•) Shielded cable F(•) 4-pole connector M12 x 1 | Z (•) 4-pole connector microsolenoid M(•) 7 pole connector 6 pole connector (•) With this kind of connector is available the "Autozero" function only. For special request contact Gefran. Ex.: TPSADA - N - 1 - P - B35D - T - V Pressure transmitter TPSADA digital autozero & span, 0...10Vdc output signal, G 1/4 male pressure connection, 7 pole connector, 0...350 bar measurement range, ± 0.1% FSO accuracy, 1msec response time. Sensors are manufactured in compliance with: - EMC 2014/30/EU compatibility directive - RoHS 2011/65/EU directive Electrical installation requirements and Conformity certificate are available on our web site: www.gefran.com

GEFRAN spa reserves the right to make any kind of design or functional modification at any moment without prior notice



via Sebina, 74 25050 PROVAGLIO D'ISEO (BS) - ITALIA tel. 0309888.1 - fax. 0309839063 Internet: http://www.gefran.com

